

# Current Utilization of Continuous End-Tidal Carbon Dioxide Monitoring in Pediatric Emergency Departments

Melissa L. Langhan, MD and Lei Chen, MD

**Objective:** End-tidal carbon dioxide (ETCO<sub>2</sub>) monitoring has numerous clinical applications in the emergency setting. This study was designed to explore the current availability and utilization patterns for continuous ETCO<sub>2</sub> monitoring in pediatric emergency departments.

**Methods:** A Web-based survey was distributed to directors of all accredited pediatric emergency medicine fellowship programs in the United States and Canada.

**Results:** Eighty-one percent of directors completed this survey. Eighty-eight percent had access to ETCO<sub>2</sub> monitoring for intubated patients and 53% for nonintubated patients. Seventy-nine percent of respondents used ETCO<sub>2</sub> monitoring "always" or "often" for endotracheal tube confirmation. Only 20% of respondents used ETCO<sub>2</sub> monitoring "always" or "often" for moderate sedation, 16% for trauma, and 6% for acid-base disturbances. One hundred percent of respondents who used ETCO<sub>2</sub> monitoring felt that it was easy to use. The most common reason for not using ETCO<sub>2</sub> monitoring was lack of equipment (65%).

**Conclusions:** ETCO<sub>2</sub> monitoring is widely available, yet underutilized, for spontaneously breathing patients in pediatric emergency departments.

**Key Words:** end-tidal carbon dioxide, monitoring

Capnography is the measurement of the exhaled carbon dioxide (CO<sub>2</sub>) during inspiration and expiration. A capnogram is the wave form representing the concentration of CO<sub>2</sub> over time. Capnometry is the numeric value of maximum exhaled CO<sub>2</sub> in the respiratory cycle, also known as the end-tidal carbon dioxide (ETCO<sub>2</sub>). ETCO<sub>2</sub> represents the alveolar concentration of carbon dioxide and has been shown to closely approximate arterial CO<sub>2</sub> concentrations.<sup>1</sup> Recent advances have made it possible to easily measure exhaled CO<sub>2</sub> in intubated patients, through an inline attachment to the endotracheal tube, as well as in patients that are spontaneously breathing through a nasal cannula. Smaller portable devices make it convenient to observe exhaled CO<sub>2</sub> levels at the bedside in a noninvasive manner, which is well tolerated by children.

Capnography and ETCO<sub>2</sub> are objective markers of ventilatory status. In addition, this monitoring tool offers a noninvasive method of measuring acid-base status and cardiac output. Current uses in emergency departments include confirmation of endotracheal intubation, monitoring the ventilatory status of intubated patients, and monitoring non-intubated patients undergoing moderate sedation.<sup>2</sup> Recent studies have also used it to measure metabolic disturbances.<sup>3,4</sup>

It has been shown, however, that continuous ETCO<sub>2</sub> monitoring is underutilized by physicians in pediatric and adult intensive care units abroad and physicians in adult emergency departments and emergency medical service personnel in the United States.<sup>5-8</sup> This is the first study to investigate how ETCO<sub>2</sub> monitoring is being used by pediatric emergency medicine (PEM) physicians. Our objective was to determine the current availability and utilization of continuous ETCO<sub>2</sub> monitoring in pediatric emergency departments (PED).

## METHODS

Between June and July 2006, a Web-based survey was distributed to the directors of all accredited PEM fellowship programs in the United States and Canada. Respondents were queried about the type of facility in which they practiced, the availability of continuous ETCO<sub>2</sub> monitoring for intubated and nonintubated patients, the perceived strengths and weaknesses of the device, and interest in learning more about ETCO<sub>2</sub> monitoring and its applications. A 4-point Likert scale was used to assess current utilization of continuous ETCO<sub>2</sub> monitoring for various clinical applications, rated as "always," "often," "sometimes," and "never."

The survey was distributed up to 3 times to ensure maximum participation. Data were analyzed with SPSS 14.0 (SPSS Inc, Chicago, IL). Descriptive analyses and chi-square analysis were performed. This study was approved by the Yale University Human Investigation Committee.

## RESULTS

Forty-three (74%) of 58 PEM fellowship directors completed the survey. Seventy-four percent stated that their primary institution was a stand-alone children's hospital, and 23% of respondents had a separate PED in a general hospital. One respondent described their facility as a separate PED adjacent to an adult emergency department. Thirty-eight respondents (88%) had access to continuous ETCO<sub>2</sub> monitoring for intubated patients and 23 respondents (53%) for nonintubated patients. Two respondents were unsure if monitoring was available for nonintubated patients. Five

Department of Pediatrics, Section of Emergency Medicine, Yale University School of Medicine, New Haven, CT.

Address correspondence and reprint requests to Melissa L. Langhan, MD, 840 Howard Avenue, Yale-New Haven Children's Hospital, New Haven, CT 06520. E-mail: Melissa.langhan@yale.edu.

Copyright © 2008 by Lippincott Williams & Wilkins  
ISSN: 0749-5161/08/2404-0211

**TABLE 1.** Current Utilization of Continuous ETCO<sub>2</sub> Monitoring in Clinical Situations (N = 38)

Clinical Situation	Always	Often	Sometimes	Never
Endotracheal tube confirmation	69%	10%	10%	10%
Moderate sedation	14%	6%	19%	61%
Trauma	8%	8%	38%	46%
Acid-base disturbances (eg, diabetic ketoacidosis, gastroenteritis)	3%	3%	22%	72%
Cardiopulmonary resuscitation	54%	8%	19%	19%

respondents did not have access to ETCO<sub>2</sub> monitoring for either intubated or nonintubated patients.

If continuous ETCO<sub>2</sub> monitoring was available, fellowship directors were asked about its current utilization in several clinical situations; these results are listed in Table 1. There was a significant difference in the frequency of utilization of ETCO<sub>2</sub> monitoring for endotracheal tube confirmation and cardiopulmonary resuscitation between respondents from stand-alone children's hospital and PEDs within general hospitals ( $P = 0.001$ ).

Of the institutions that currently use continuous ETCO<sub>2</sub> monitoring, 34 (100%) felt that it is easy to use, 33 (97%) felt that the results are accurate, and 31 (91%) felt that the results are easy to interpret. Furthermore, 25 respondents (74%) felt knowledgeable regarding current ETCO<sub>2</sub> literature and 19 (59%) felt that the equipment for continuous ETCO<sub>2</sub> monitoring was inexpensive.

Of the 18 respondents who did not routinely use continuous ETCO<sub>2</sub> monitoring in the PED, their primary reason was lack of equipment (65%). Other reasons included lack of familiarity (33%), lack of perceived need (22%), perception of poor evidence for use in the medical literature (22%), perceived inaccurate measurements (11%), expense (11%), and difficulty with use (6%).

Thirty-one respondents (78%) were interested in learning more about continuous ETCO<sub>2</sub> monitoring and its applications. The majority (45%) would prefer receiving literature, whereas 35% desired online information, 33% would attend a workshop, and 33% would view an on-site demonstration.

## DISCUSSION

This study demonstrated that continuous ETCO<sub>2</sub> monitoring is available in most PEDs for both intubated and nonintubated patients. Similarly, most PEM fellowship directors have positive opinions regarding the technical aspects of this device. Despite these views, however, ETCO<sub>2</sub> monitoring is not being used in many common medical and traumatic situations that occur in the PED.

ETCO<sub>2</sub> has been studied in a variety of clinical settings. Several studies have demonstrated the superiority of capnometry and capnography over auscultation in the detection of esophageal intubations in patients who were

emergently intubated.<sup>9,10</sup> As a continuous monitor of ventilatory status and a marker of cardiac output, this device is useful in several emergency department situations in addition to confirmation of endotracheal intubations. These conditions where noninvasive monitoring of ETCO<sub>2</sub> may be beneficial include moderate sedation, status epilepticus, sepsis, trauma, and cardiopulmonary resuscitation.<sup>11-14</sup>

The most frequently studied application of capnography for nonintubated patients is in the area of sedation and analgesia. It has been proposed that ETCO<sub>2</sub> monitoring can improve patient safety. It is a valuable marker of hypoventilation and apnea. Research has demonstrated that ETCO<sub>2</sub> can detect trends toward respiratory depression with much greater sensitivity than either pulse oximetry or clinical observations, which are the current standards of care.<sup>15,16</sup> Furthermore, ETCO<sub>2</sub> monitoring has shown objective differences in the degree of hypercapnia based on the type of medications used for sedation. Children who receive propofol or a combination of midazolam and an opiate for procedural sedation have been shown to have higher ETCO<sub>2</sub> levels than those with other medications including ketamine, ketamine with midazolam, or midazolam alone.<sup>12,16,17</sup>

Exhaled CO<sub>2</sub> can be a marker of metabolic disturbances. Therefore, it is also helpful in managing conditions such as diabetic ketoacidosis and dehydration.<sup>3,4</sup> Fearon and Steele<sup>3</sup> demonstrated a linear relationship between ETCO<sub>2</sub> measurements and serum bicarbonate levels in children with diabetes. ETCO<sub>2</sub> was shown to correlate with the severity of metabolic acidosis and thus acted as a predictor of diabetic ketoacidosis in comparison to hyperglycemia. Similarly, Nagler et al demonstrated a linear correlation between ETCO<sub>2</sub> measurements and serum bicarbonate levels in children with dehydration from gastroenteritis. Children with lower ETCO<sub>2</sub> levels were more likely to be severely dehydrated and require treatment with intravenous fluids than those who received oral rehydration and were assessed as being less dehydrated. In addition, children who had unscheduled return visits had lower ETCO<sub>2</sub> levels than those who did not seek further medical evaluation.<sup>4</sup>

This study is limited by its design and sample size. There are a small number of accredited PEM fellowship programs across the United States and Canada. Similarly, we are relying on the experiences of one physician within each institution to represent the group as a whole. On the other hand, PEM fellowship programs are usually based in academic settings. It is likely that these represent the forefront of utilization of new technologies. One may argue that the availability and knowledge of this technology would be greater than that of physicians in other nonacademic and community hospitals.

Although continuous ETCO<sub>2</sub> monitoring is widely used in PEDs for intubated patients, our survey results indicated that there was far less utilization of this technology for children who are spontaneously breathing, despite a fair body of evidence supporting its use. Further development and dissemination of training curriculum, such as workshops at future fellows' conferences, may expand the utilization of this noninvasive diagnostic modality for the benefit of patients in the PED.

## REFERENCES

1. Casati A, Gallioli G, Scandroglio M, et al. Accuracy of end-tidal carbon dioxide monitoring using the NBP-751 microstream capnometer. A study in intubated ventilated and spontaneously breathing nonintubated patients. *Eur J Anesth*. 2000;17:622–626.
2. Bhende MS. End tidal carbon dioxide monitoring in pediatrics—clinical applications. *J Postgrad Med*. 2001;47:215–218.
3. Fearon DM, Steele DW. End-tidal carbon dioxide predicts the presence and severity of acidosis in children with diabetes. *Acad Emerg Med*. 2002;9:1373–1378.
4. Nagler J, Wright RO, Krauss B. End-tidal carbon dioxide as a measure of acidosis among children with gastroenteritis. *Pediatrics*. 2006;118:260–267.
5. Erasmus PD. The use of end-tidal carbon dioxide monitoring to confirm endotracheal tube placement in adult and paediatric intensive care units in Australia and New Zealand. *Anaesth Intensive Care*. 2004;32:672–675.
6. Delorio NM. Continuous end-tidal carbon dioxide monitoring for confirmation of endotracheal tube placement is neither widely available nor consistently applied by emergency physicians. *Emerg Med J*. 2005;22:490–493.
7. Cumming C, McFadzean J. A survey of the use of capnography for the confirmation of correct placement of tracheal tubes in pediatric intensive care units in the UK. *Pediatr Anesth*. 2005;15:591–596.
8. Bair AE, Smith D, Lichty L. Intubation confirmation techniques associated with unrecognized non-tracheal intubations by pre-hospital providers. *J Emerg Med*. 2004;28:403–407.
9. Grmec S. Comparison of three different methods to confirm tracheal tube placement in emergency intubation. *Intensive Care Med*. 2002;28:701–704.
10. Silvestri S, Ralls GA, Krauss B, et al. The effectiveness of out-of-hospital use of continuous end-tidal carbon dioxide monitoring on the rate of unrecognized misplaced intubations within a regional emergency medical services system. *Ann Emerg Med*. 2005;45:497–503.
11. Abramo TJ, Wiebe RA, Scott S, et al. Noninvasive capnometry monitoring for respiratory status during pediatric seizures. *Crit Care Med*. 1997;25:1242–1246.
12. Yildizdas D, Yapicioglu H, Yilmaz HL. The value of capnography during sedation or sedation/analgesia in pediatric minor procedures. *Pediatr Emerg Care*. 2004;20:162–165.
13. Gazmuri RJ, Kube E. Capnography during cardiac resuscitation: a clue on mechanisms and a guide to interventions. *Critical Care*. 2003;7:411–412.
14. Deakin CD, Sado DM, Coats TJ, et al. Prehospital end-tidal carbon dioxide concentration and outcome in major trauma. *J Trauma*. 2004;57:65–68.
15. Burton JH, Harrah JD, Germann CA, et al. Does end-tidal carbon dioxide monitoring detect respiratory events prior to current sedation monitoring practices? *Acad Emerg Med*. 2006;13:500–504.
16. Deitch K, Chudnofsky CR, Dominici P. The utility of supplemental oxygen during emergency department procedural sedation and analgesia with midazolam and fentanyl: a randomized, controlled trial. *Ann Emerg Med*. 2007;49:1–8.
17. McQuillen KK, Steele DW. Capnography during sedation/analgesia in the pediatric emergency department. *Pediatr Emerg Care*. 2000;16:401–404.